

THOMSON REUTERS

**WESTLAW California Code of Regulations**[Home Table of Contents](#)**§ 75062. Maintenance and Operation.**

22 CA ADC § 75062

**BARCLAYS OFFICIAL CALIFORNIA CODE OF REGULATIONS**Barclays Official California Code of Regulations Currentness

Title 22. Social Security

Division 5. Licensing and Certification of Health Facilities, Home Health Agencies, Clinics, and Referral Agencies

Chapter 7. Primary Care Clinics

Article 7. Physical Plant (Refs &amp; Annos)

22 CCR § 75062

**§ 75062. Maintenance and Operation.**

- (a) The clinic shall be clean, sanitary and in good repair at all times.
- (b) Flashlights shall be in readiness for use at all times. Open flame light shall not be used.
- (c) Periodic inspection, testing and calibration of all electrical medical equipment shall be made in accordance with the manufacturer's specifications, but no less frequently than yearly. Records of such inspections and tests shall be retained and made available for review.
- (d) A clinic may be required to submit a report by a licensed structural engineer when the Department has determined that an evaluation of the structural condition of the clinic is necessary. Such report shall establish the existence or nonexistence of structural conditions which are hazardous to occupants and should hazardous conditions exist, such report shall provide a basis for a plan of correction.

Note: Authority cited: Sections 208(a) and 1225, Health and Safety Code. Reference: Section 1226, Health and Safety Code.

This database is current through 6/30/17 Register 2017, No. 26

22 CCR § 75062, 22 CA ADC § 75062

**END OF DOCUMENT**

© 2017 Thomson Reuters. No claim to original U.S. Government Works.