

SUBJECT: **ASSESSMENT UNIT INVENTORY**

CHANGE LOG

Published Date	Status	Section and Subsection Affected	Description of Change(s)
07/01/2023	Addition, Revision, +Deletion	Supplies; Intravenous Fluids; Approved Optional Equipment	<ul style="list-style-type: none"> • Added Manometer x 2 to required inventory • Added under Cardiac Monitor in required inventory - pulse oximeter, waveform capnography, noninvasive blood pressure monitor, transcutaneous pacing, ECG, 12 lead and transmission capable • Bag Mask device changed from 650-1000 mL to >900 mL in required inventory • Removed Aspirator from required inventory • Simple Masks moved to optional inventory • Added airflow meter with rate and volume capability to optional inventory
09/01/2023	Addition	Supplies; Approved Optional Equipment Footnotes	<ul style="list-style-type: none"> • Added toddler bag-mask ventilation masks; added macrodrip to intravenous tubing • Added "EMS Authority Approved Dressing Only" to hemostatic dressing, Tranexamic Acid (TXA) 1gm, Normal saline 50 or 100 mL bag • Added information on TXA in footnotes