

SUBJECT: **RESPIRATORY CARE PRACTITIONER STAFFED
SPECIALTY CARE TRANSPORT UNIT INVENTORY** REFERENCE NO. 713

CHANGE LOG

Published Date	Status	Section and Subsection Affected	Description of Change(s)
07/01/2023	Addition + Revision	Supplies; Approved Optional Equipment	<ul style="list-style-type: none">• Added Manometer x 2 to required inventory• Bag Mask device changed from 650-1000 mL to >900 mL• Airflow meter with rate and volume capacity added to optional inventory
09/01/2023	Revision	Supplies	<ul style="list-style-type: none">• Revised verbiage for Continuous Positive Airway Pressure (CPAP)